



# Auscultation Manikin Update Package

**KERi**™ — LF04024U

**GERI™** — LF04004U

**Instruction Manual** 





Congratulations for choosing the **GERI™/KERI™** Auscultation Update Package. This simple design makes the manikin easy to use and care for. A diagnosis of the condition selected by the instructor can be made by comparing the variation in sounds occurring at different sites — just like a real person! Lung sounds can be detected at five anterior and 12 posterior locations, and students can practice auscultation at six anterior heart sites. Before using this manikin with your students, please familiarize yourself with this manual.

### **List of Components**

- GERi™/KERi™ Manikin Update
- SmartScope<sup>™</sup>
- Remote Control



Figure 1

#### **General Instructions for Use**

The **GERi™/KERi™** Update comes ready to be attached to your existing manikin. To do this, you will need to use the arms, legs, and head of your existing manikin. The arms can be

removed by bending the elbows backward approximately 90° to align the keyholes at the shoulders. (See figure 1.) Then pull outward, away from the shoulder of the manikin. Reverse this procedure to attach the arms to the new GERi™/KERi™ Auscultation Torso.



Figure 2

The legs detach from the body at the hips by bending the legs backward approximately 150° so that the feet are near the shoulders and the keyholes are aligned. (See figure 2.) Then pull outward, away from the body. Reverse this procedure to attach to the new torso.



Figure 3

The head may be removed by rotating it 180° backward to align the keyholes, and then pulling up. The internal stomach reservoir will come out of the body through the neck opening. Reverse this procedure to attach the head to the new torso.

## Gastrostomy Care Lavage and Gavage

A flanged hole simulating an abdominal incision for the insertion of a feeding tube is included on the upper torso for performing lavage and gavage. Inside the upper torso is a reservoir (bag) with a maximum capacity of 500 cc, which is attached to the gastrostomy opening with a two-part coupler. The bag is removed by pushing in on the clear L-shaped button and pulling it straight away. To attach, simply push the coupler body (with bag attached) onto the coupler insert (part with a black O-ring) that is protruding from inside the torso. You will hear a slight snap when the connection is complete. With extensive use, these two coupler parts may no longer snap, so the clear L-shaped button will have to be manually pulled all the way out after connecting the two parts. Test the connection by pulling on the coupler body to ensure that it is locked. This gastrostomy feature is designed for use with a 16 French feeding tube. It is recommended that the tip of the feeding tube be lubricated before inserting. Make sure the reservoir is straight and flat before attempting to simulate actual feeding with a liquid.

#### **About the Auscultation Simulator**

The **GERI™/KERI™** Auscultation Manikin duplicates heart and lung conditions selected by the instructor by wireless remote control. The student should palpate to identify the correct auscultation sites. They will hear different heart and lung sounds as the SmartScope™ is moved from site to site.

The simulator has six heart sites and seven lung sound sites on the anterior surface, with 10 locations on the posterior surface and midaxillary sites. The remote control can select from 12 different heart conditions, as well as 16 lung conditions. The instructor can select any condition easily, enabling the student to compare sounds and make a diagnosis.

The remote control does not have to be pointed directly at the manikin or stethoscope to operate. One remote control will operate multiple sets of SmartScopes™ and manikins simultaneously. Great for group instruction. The range of the remote control is up to 100 feet.

#### General Instructions for Use

To begin using the Auscultation Trainer, press the red power button. This turns on the remote control and sends a signal to activate the stethoscope as well. After the unit is activated, the display will be in the "status" mode, displaying the current menu settings for the heart and lung conditions.

To select a new condition, press either the heart or lung button. This will put the display into the menu mode. The user can select a condition by either using the number buttons, or by viewing the conditions in sequence using the scroll button. After the condition is selected, press the enter button to activate.

#### The remote control can be programmed to these heart conditions:

- 01 Normal
- 02 Aortic regurgitation
- 03 Pulmonary stenosis
- 04 Mitral stenosis
- 05 Holosystolic
- 06 Mid-systolic
- 07 S3 Gallop
- 08 S4 Gallop
- 09 Systolic click
- 10 Atrial septal defect
- 11 PDA
- 12 VSD

### The remote control can be programmed to these lung conditions:

- 01 Normal Lung
- 02 Normal Vesicular
- 03 Wheezes
- 04 Mono wheeze
- 05 Fine crackle
- 06 Coarse crackle
- 07 Ronchi crackle
- 08 Stridor
- 09 Cavernous
- 10 Bronchovesicular
- 11 Bronchial
- 12 Pulmonary edema
- 13 Infant
- 14 Friction rub
- 15 Egophony
- 16 Pectoriloquy

Consult the enclosed laminated card for lung and heart sound site locations.

To listen to the selected sounds, place the earpieces in ears, angled in a forward position. Place the diaphragm of the SmartScope™over one of the appropriate sites on the manikin (see diagram on back).

As an alternative, an Amplifier/Speaker System (LF01189U) can be used, allowing an entire classroom to hear at the same time. To connect the speaker, plug the speaker cord into the speaker jack on the top of the Smart Scope™ box. When a speaker is connected to the SmartScope™, the earpieces will not work. Place the diaphragm of the SmartScope™ over the appropriate sites on the manikin.

**Note:** The remote has a battery saver that shuts the unit down after eight minutes if the remote is left on the same setting. To prevent it from shutting down, select a different heart or lung sound within the eight-minute period of time.

### **Available Supplies**

**LF01144U** Additional SmartScope™

**LF01148U** Replacement Remote Control

LF01189U Amplifier/Speaker System

# GERI<sup>TM</sup>/KERI<sup>TM</sup> THE NURSING SKILLS MANIKIN Component Replacement Parts

